## **Installation for Safe and Efficient Operation**

 $\triangleright$ 

The oscilloscope will operate to its specifications if the operating environment is maintained within the following parameters:

**Operating Environment** 

3

- Temperature ......5 to 40 °C (41 to 104 °F) rated.
- Humidity......Maximum relative humidity 80 % RH (non-condensing) for temperatures up to 31 °C decreasing linearly to 50 % relative humidity at 40 °C
- Altitude.....< 2000 m (6560 ft)</p>

The oscilloscope has been qualified to the following EN61010-1 category:

- Protection Class ...... I
- Installation (Overvoltage) Category......II

Safety Symbols

Where the following symbols or indications appear on the instrument's front or rear panels, or elsewhere in this manual, they alert the user to an aspect of safety.

Symbol	Meaning
Â	CAUTION: Refer to accompanying documents (for Safety- related information). See elsewhere in this manual wherever the symbol is present, as indicated in the Table of Contents.
Â	CAUTION: Risk of electric shock.
	On (Supply).

3–1

Symbol	Meaning	
$\bigcirc$	Off (Supply)	
Ļ	Earth (Ground) Terminal	
	Protective Conductor Terminal	
$\rightarrow$	Chassis Terminal	
↓	Earth (Ground) Terminal on BNC Connectors	
WARNING	Denotes a hazard. If a WARNING is indicated on the instrument, do not proceed until its conditions are understood and met.	
WARNING	Any use of this instrument in a manner not specified by the manufacturer may impair the instrument's safety protection. The oscilloscope has <i>not</i> been designed to make direct measurements on the human body. Users who connect a LeCroy oscilloscope directly to a person do so at their own risk. Use only indoors.	
Power Requirements	The oscilloscope operates from a 115 V (90 to 132 V) or 220 V (180 to 250 V) AC power source at 45 Hz to 66 Hz.	

3–2

	No voltage selection is required, since the instrument automatically adapts to the line voltage present.		
Fuses	The oscilloscope's power supply is protected against short- overload by means of two "T"-rated fuses of type according model:		
	> 6.3 A/250 V AC	9344C, 9350C, 9354C, 9370C, 9374C, 9384C Series	
	> 5 A/250 V AC	9304C, 9310C, 9314C Series.	
	The fuses are located all before inspecting or rep small screwdriver under continued fire protection specified type and rating	bove the mains plug. Disconnect the power cord placing a fuse. Open the fuse box by inserting a er the plastic cover and prying it open. For at all line voltages, replace only with fuses of the (see above).	
Ground	The oscilloscope has been designed to operate from a single-phase power source, with one of the current-carrying conductors (neutra conductor) at ground (earth) potential. Maintain the ground line to avoid an electric shock. None of the current-carrying conductors may exceed 250 V rms with respect to ground potential. The oscilloscope is provided with a three-wire electrical cord containing a three- terminal polarized plug for mains voltage and safety ground connection. The plug's ground terminal is connected directly to the frame of the unit. For adequate protection against electrical hazard this plug must be inserted into a mating outlet containing a safety ground contact.		
Cleaning and Maintenance	Maintenance and repairs should be carried out exclusively by a LeCrot technician ( <i>see Chapter 1</i> ). Cleaning should be limited to the exterior of the instrument only, using a damp, soft cloth. Do not use chemicals of abrasive elements. Under no circumstances should moisture be allower to penetrate the oscilloscope. To avoid electric shocks, disconnect the instrument from the power supply before cleaning.		
	Risk of electrical shock: No user-serviceable parts inside. Leave repair to qualified personnel.		
Power On	Connect the oscilloscop the power switch locat switched on, auto-calibra ADCs and memories i approximately 10 secon screen.	e to the power outlet and switch it on by pressing ed on the rear panel. After the instrument is ation is performed and a test of the oscilloscope's s carried out. The full testing procedure takes ids, after which time a display will appear on the	

