

Manual Supplement

Manual Title:	Y5020	Supplement Issue:	1
Part Number:	516591	Issue Date:	4/81
Print Date:	September 1979	Page Count:	1
Revision/Date:			

This supplement contains information necessary to ensure the accuracy of the above manual. Enter the corrections in the manual if either one of the following conditions exist:

1. The revision letter stamped on the indicated PCA is equal to or higher than that given with each change.
2. No revision letter is indicated at the beginning of the change.

Change #1

On page 1, Table 1, make changes to meet required specifications:

Change:	ACCURACY OF CALIBRATION VALUE	$\pm .01\%$
To:	ACCURACY OF CALIBRATION VALUE	$\pm .01\%$ at dc
Delete:	DC to 1 kHz	$\pm .025\%$
Delete:	1 kHz to 5 kHz	$\pm .035\%$
Add:	AC - DC DIFFERENCE	$\pm (.015\% + .012\% \times f) f$ (kHz)
Change:	STABILITY OF CALIBRATION VALUE	(30 ppm per 6 months)
To:	STABILITY OF CALIBRATION VALUE	Less than .002% change in resistance in six months, .003% for 1 year